

9-05-2006

České dráhy, a.s.

(Czech Railways)

Železniční zdravotnictví odštěpný závod

(Railway Health Service)

Železniční nemocnice Praha

(Railway Hospital Prague)

Prague 2, Italská 37, post code 121 43

Company reg.nr.: 70 99 42 26, tax reg.nr.: CZ70 99 42 26

Incorporated at Municipal Court in Prague, Section B, insert 8039

PROTOCOL ON OSMOLALITY ASSESMENT

On 15 March 2006 in the laboratory of Railway Hospital Prague there was set a value of osmolality at the submitted sample of an isotonic drink ADAPT SPORT from the orderer, GENESYS EUROPE s.r.o. company.

Evaluation: osmolality value is 290 mosmol/kg

Assesment made by osmometer GONOTEC 300.

Calibration made at the same time, caliber SONO-OSMO, producer Pliva-Lachema Brno.

In Prague, dated 15 March 2006

RNDr J. Sýkora - *signature*

Department of clinical biochemistry

Railway Hospital Prague

Stamp

Stamp: Verification – vidimus clause

*I verify the copy of 1 sheet is verbatim in accordance with
the original copy of 1 sheet , from which it has been made.*

In Ostrava, dated 27 June 2006

Notary public JUDr. Miluše Rabasová - signature

Seal: JUDr. Miluše Rabasová

Notary public in Ostrava